



## Certificate of Analysis


|                 |                                  |
|-----------------|----------------------------------|
| Product Name    | : BMPro-V Medium (Phenol Red(-)) |
| Cat. No.        | : 2501P10                        |
| Size            | : Bottle 1000 mL                 |
| Lot No.         | : 0126260D                       |
| Expiration Date | : 2026.10                        |

| EXAMINATIONS                  |          | SPECIFICATIONS          | RESULTS       |
|-------------------------------|----------|-------------------------|---------------|
| Appearance                    |          | Colorless, Clear Liquid | PASS          |
| pH *1                         |          | 7.4 ± 0.2               | 7.4           |
| Osmotic Press. *2             |          | 285 ± 10 mOsm           | 285 mOsm      |
| Endotoxin *2                  |          | < 0.25 EU/mL            | < 0.003 EU/mL |
| Sterility Test *2             | Bacteria | Negative                | PASS          |
|                               | Fungi    | Negative                | PASS          |
| Cell Culture Test (Vero cell) |          | PASS                    | PASS          |

\*1 After incubate in a 5%CO<sub>2</sub> incubator.

\*2 JP

We prove that this product adapted to above specifications.

  
Satori Ishiya  
QC director

2026.02.12

Acceptance Date

Note to User: Culture media are reagent for research use, do not use for clinical therapy and diagnostic procedure, and strictly prohibited an injection into a body directly.

The followings are registered trademarks of Cell Science & Technology Institute Inc., (CSTI).

CSTI® BMPro®